

CERTIFICATE OF ANALYSIS No.: 2022-10152

CLIENT

CIITECH Ltd, 2 Athenaeum Road
GB-N20 9AE London, United Kingdom



SAMPLE *

Provacan CBD Oil Drops 300mg / 3% Full Spectrum,
10ml

Sample condition: SUITABLE
Sample ID: 2242037
Sample type: Viscous liquid
Batch No.: * DR03122286A

Work order: 2022-107014
Analysis ID: 2022_241c
Method ID: PHL_RPC_12C
Method SOP: MET-LAB-003-02

Sample received: 20/10/2022
Start of analysis: 21/10/2022
End of analysis: 24/10/2022
Analyst: Janja Ahej

* Information provided by the client.

CANNABINOID TRACE ANALYSIS

	Concentration [% w/w]	Expanded uncertainty [% w/w]	LOQ [% w/w]	Graphic presentation of relative cannabinoid concentration
CBDV - Cannabidivarin	0.0154	0.0035	0.00300	
CBDA - Cannabidiolic acid	0.118	0.020	0.00300	
CBGA - Cannabigerolic acid	< LOQ	n/a	0.00300	
CBG - Cannabigerol	0.47	0.12	0.00300	
CBD - Cannabidiol	3.06	0.15	0.00300	
THCV - Tetrahydrocannabivarin	< LOQ	n/a	0.00300	
CBN - Cannabinol	0.0100	0.0022	0.00300	
Δ⁹-THC - Δ-9-Tetrahydrocannabinol	< LOQ	n/a	0.00300	
Δ⁸-THC - Δ-8-Tetrahydrocannabinol	< LOQ	n/a	0.00300	
CBL - Cannabicyclol	< LOQ	n/a	0.03000	
CBC - Cannabichromene	0.0189	0.0042	0.00300	
Δ⁹-THCA - Δ-9-Tetrahydrocannabinolic acid	< LOQ	n/a	0.00300	
CBE - Cannabielsoin	0.0159 #	0.0045	0.00300	
CBNV - Cannabivarin	< LOQ #	n/a	0.00300	
CBCA - Cannabichromenic acid	< LOQ #	n/a	0.03000	
CBT - Cannabicitran	0.0228 #	0.0050	0.00300	

Units and abbreviations: % w/w = weight percent, LOQ = the limit of quantitation, ND = not detected, n/a = not available.

The results given herein apply only to the sample as received. **Expanded Uncertainty** was calculated using coverage factor $k = 2$, corresponding to a double standard uncertainty and characterizes the interval value in which it is possible to expect the real value with a probability of 95%. This is stated according to the ISO/IEC Guide 98-3.

Total or partial reproduction of this document is not allowed without the permit from PharmaHemp d.o.o. The document does not substitute any other legal document.

Date issued:

24/10/2022

Approved by:

mag. Marko Dragan
Analytical Laboratory Manager

Authorized by:

dr. Boštjan Jančar
Chief Technology Officer

End of Certificate